

PLTP Activity Assay Kit

Cat. No. Kit-0707 **Lot. No.** (See product label)

SPECIFICATION

Product Overview

Simple procedure; takes ~40 minutes
 Fast and convenient
 Supplied with 100 positive control assays and all other necessary reagents for screening of PLTP inhibitors.

Description

Phospholipid transfer protein (PLTP) plays an important role in transferring phospholipids between HDL molecules to modulate HDL size and composition and controlling plasma HDL levels. PLTP also transfers phospholipids from lipoproteins to HDL. PLTP plays a key role in reverse cholesterol transport and may promote atherosclerosis. PLTP is, therefore, considered a promising target for pharmacological intervention. PLTP Activity Fluorometric Assay Kit II uses a self-quenched fluorescent phospholipid that can be measured when transferred to an acceptor molecule. The fluorometric intensity is directly proportional to the amount of phospholipid transferred. Rabbit serum is provided as a positive control. This Assay Kit, in addition to measuring activity in serum, is also suitable for testing activity of recombinant protein.

Usage

This product is furnished for LABORATORY RESEARCH USE ONLY. Not for diagnostic or therapeutic use.

Size 100 assays

Kit Components

PLTP Donor Molecule
 PLTP Acceptor Molecule

 Tel: 1-631-559-9269 1-516-512-3133

 Email: info@creative-biomart.com  Fax: 1-631-938-8127

 45-1 Ramsey Road, Shirley, NY 11967, USA

	PLTP Assay Buffer (10X) Positive Control (Rabbit Serum)
Target Species	Mammalian
Detection method	Fluorometer
Compatible Sample Types	Cell Culture Supernatants Plasma Serum Urine
Preparation	Warm Assay Buffer to room temperature before use. Briefly centrifuge small vials prior to opening. All kit components are supplied as ready to be used. Keep on ice while in use.
Assay Protocol	<p>1. Standard Curve Preparation: Make serial dilutions of the Donor Molecule in 100% isopropanol. Dilute Donor Molecule 100 times by adding 10 µof Donor Molecule to 990 µof 100% isopropanol. Dilute further by adding 100 µof 100 times diluted donor molecule into 900 µof 100% isopropanol and label as T5. Label four Eppendorf tubes as T4, T3, T2 and T1 respectively. Aliquot 250 µof isopropanol into each tube. Add 250 µfrom T5 into T4 and mix. Transfer 250 µfrom T4 into T3 and mix, and continue for T2 and T1. Add 200 µfrom each tube into a series of wells in 96-well plate to generate 0.375, 0.75, 1.5, 3.0 , 6.0 pmol Donor Molecule Standard. Use 200 µof 100% isopropanol as 0 pmol (b lank) Standard. Measure Fluorescence (Ex/Em = 465/535 nm). To save time, Standard Curve can be made during sample incubation.</p> <p>2. Sample Preparation: Collect plasma or serum by standard methods and keep on ice for immediate use or store at -80°C. To measure sample's PLTP activity, prepare 200 µmix containing:</p> <p>Donor Molecule 2 µ</p>

 Tel: 1-631-559-9269 1-516-512-3133

 Email: info@creative-biomart.com  Fax: 1-631-938-8127

 45-1 Ramsey Road, Shirley, NY 11967, USA

Acceptor Molecule 5 μ

Sample (plasma or serum) 1-8 μ

PLTP Assay Buffer To a total of 200 μ

For positive control, dilute rabbit serum 10 times in Assay Buffer and add 8 μ of diluted Positive Control instead of your sample in desired well (s). For the reagent background control, don't add the PLTP source i.e. plasma, serum, or recombinant protein to the reaction and make up the volume with PLTP Assay Buffer.

Notes:

- a. For unknown samples, we suggest doing a pilot experiment by testing several amounts to ensure the readings are within the Standard Curve range.
- b. Using higher than recommended amounts of plasma or serum will inhibit the signal (>2 μ undiluted). Typically diluting human or rabbit plasma 10 times and measuring 2-10 μ will give a signal within range of the Standard Curve.
3. Measurement: Pre-incubate at 37°C for 10 min. protected from light to stabilize the signal. Measure fluorescence (Ex/Em = 465/535nm) kinetically for 1-3 hr in a microplate reader at 37°C.

Note:

Incubation time depends on sample's PLTP activity. We recommend measuring fluorescence in kinetic mode and choosing two time points (T1 and T2) in the linear range to calculate the PLTP activity of the samples. The Standard Curve can be read in the end point mode.

4. Calculation: Subtract 0 Standard reading from all Standard readings. Plot the Donor Molecule Standard curve. Subtract reagent background control reading from sample reading.

$$RFU1 = RFU1S - RFU1B \quad RFU2 = RFU2S - RFU2B$$

Where: RFU1S & RFU2S is the sample reading at time T1 and T2 respectively.

RFU1B & RFU2B is the reagent background control reading at time T1 and T2 respectively.

Calculate the PLTP activity of the samples $\Delta RFU = RFU2 - RFU1$. Apply the ΔRFU to

 Tel: 1-631-559-9269 1-516-512-3133

 Email: info@creative-biomart.com  Fax: 1-631-938-8127

 45-1 Ramsey Road, Shirley, NY 11967, USA



the Standard Curve to get B pmol of phospholipids transferred by PLTP during the reaction time ($\Delta T = T_2 - T_1$). Calculate sample's PLTP activity by using the following equation:

Sample PLTP Activity (A) = $B / (\Delta T \times V) \times D = \text{pmol/ml/hr} = \text{mU/ml}$


Where: B is amount of Phospholipid from Standard Curve (pmol)

V is sample volume added into the reaction well (ml)


ΔT is reaction time (hr)

D is sample Dilution factor

Unit Definition: One unit of PLTP is the amount of protein that will transfer 1.0 nmol of donor molecule per hr at 37°C.

 Tel: 1-631-559-9269 1-516-512-3133

 Email: info@creative-biomart.com  Fax: 1-631-938-8127

 45-1 Ramsey Road, Shirley, NY 11967, USA