

Intracellular Fluorimetric Hydrogen Peroxide Assay Kit (Blue, Optimized for Flow Cytometry)

Cat. No. Kit-0988 Lot. No. (See product label)

SPECIFICATION

Product Overview	This Intracellular Fluorimetric Hydrogen Peroxide Assay Kit uses our unique Blue peroxide sensor to quantify hydrogen peroxide in live cells. Blue peroxide sensor is cell-permeable, and generates blue fluorescence when it reacts with hydrogen peroxide. This kit provides a sensitive tool to monitor hydrogen peroxide level in living cells, and it is optimized to be used in flow cytometry.
Description	Hydrogen peroxide (H ₂ O ₂) is a reactive oxygen metabolic by-product that serves as a key regulator for a number of oxidative stress-related states. It is involved in many biological events that are linked to asthma, atherosclerosis, diabetic vasculopathy, osteoporosis, a number of neurodegenerative diseases and Down's syndrome. The measurement of this reactive species is helpful for determining how oxidative stress modulates various intracellular pathways.
Storage	Keep in freezer and avoid exposure to light.
Size	100 Tests
Kit Components	Component A: Blue peroxide sensor 1 vial Component B: DMSO 1 vial (200 µL)
Preparation	1. Prepare cells: For each sample, prepare cells in 0.5 mL growth medium or buffer of your choice at a density of 5×10 ⁵ to 1×10 ⁶ cells/mL.

 Tel: 1-631-559-9269 1-516-512-3133

 Email: info@creative-biomart.com  Fax: 1-631-938-8127

 45-1 Ramsey Road, Shirley, NY 11967, USA

Note: Each cell line should be evaluated on an individual basis to determine the optimal cell density for hydrogen peroxide induction.

2. Prepare Blue peroxide sensor stock solution:

Add 100 μ of DMSO (Component B) into the vial of Blue peroxide sensor (Component A), and mix them well.

Note: 1 μ L of reconstituted Blue peroxide sensor stock solution is for 0.5 mL cells.

The stock solution should be used promptly. Any remaining solution should be aliquoted and refrozen at -20°C. Avoid repeated freeze-thaw cycles and protect from light.

3. Run the hydrogen peroxide assay:

3.1 Stain cells with Blue peroxide sensor in full medium or in your desired buffer at 37°C for 20- 30 minutes, protected from light.

3.2 Treat cells with test compounds in full medium or in your desired buffer at 37°C for desired period of time. For control samples (untreated cells), add the corresponding amount of compound buffer.

Note 1: It's recommended to treat cells in full medium. However, if tested compounds are serum sensitive, growth medium and serum factors can be aspirated away before treatment. Resuspend cells in 1X Hank's salt solution and 20 mM Hepes buffer (HHBS) or the buffer of your choice after aspiration. Alternatively, cells can be treated in serum-free media.

Note 2: We treated Jurkat cells with 100 μ M hydrogen peroxide in full medium at 37°C for 90 minutes to induce hydrogen peroxide.

3.3 Monitor the fluorescence intensity at Pacific Blue channel (Ex/Em=405/450 nm) using a flow cytometer. Gate on the cells of interest, excluding debris.

Assay Protocol

 Tel: 1-631-559-9269 1-516-512-3133

 Email: info@creative-biomart.com  Fax: 1-631-938-8127

 45-1 Ramsey Road, Shirley, NY 11967, USA