

Serum total iron binding capacity (TIBC) test kit

Cat. No. KITZ-080 **Lot. No.** (See product label)

SPECIFICATION

Product Overview	Excess iron is added to the serum to ensure that all the transferrin in the serum is bound to iron. An iron adsorbent is then added to adsorb the excess iron. The iron content is subsequently measured using a serum iron assay, and this value is referred to as the total iron-binding capacity (TIBC). By subtracting the serum iron value from the TIBC, the unsaturated iron-binding capacity (UIBC) is obtained.
Size	96 Samples
Storage	4°C
Shipping	Ice pack
Kit Components	Reagent A: Liquid, 21mL × 1 bottle, stored at 4°C. Reagent B: Powder × 1 bottle, stored at room temperature. Reagent 1: Liquid, 26mL × 1 bottle, stored at 4°C. Reagent 2: Powder × 2 vials, stored at 4°C. Before use, shake well or centrifuge to bring the powder to the bottom, then add 1.2mL of distilled water to each vial to dissolve and prepare for use. Reagent 3: Liquid, 2mL × 1 vial, stored at 4°C. Standard: Liquid, 1mL × 1 vial, stored at 4°C. Before use, dilute the standard with Reagent 3 by 50 times (i.e., take 10µL of the standard into an EP tube, then add 490µL of Reagent 3) to prepare a 2mg/L, which is 35.82 µmol/L, iron standard solution.

 Tel: 1-631-559-9269 1-516-512-3133

 Email: info@creative-biomart.com  Fax: 1-631-938-8127

 45-1 Ramsey Road, Shirley, NY 11967, USA

**Materials Required
but Not Supplied**

Microplate reader, 96-well plate, adjustable pipette, centrifuge, distilled water.

 Tel: 1-631-559-9269 1-516-512-3133

 Email: info@creative-biomart.com  Fax: 1-631-938-8127

 45-1 Ramsey Road, Shirley, NY 11967, USA