Monoamine Oxidase ELISA Detection Kit

Cat. No. KITG-043 **Lot. No.** (See product label)

SPECIFICATION

Product Overview	This kit is for research use only and not for use in medical diagnosis. This ELISA Detection Kit is designed for the quantitative detection of Monoamine Oxidase (MAO) in mouse samples using a double-antibody sandwich enzyme-linked immunosorbent assay (ELISA) method. The assay utilizes a pre-coated microplate with monoamine oxidase (MAO) antibody. Samples, standards, and HRP-labeled detection antibody are added sequentially to the coated wells. After incubation and thorough washing, the substrate TMB is added for color development. TMB is converted to blue under the catalysis of peroxidase and then to yellow upon acidification. The intensity of color is positively correlated with the concentration of monoamine oxidase (MAO) in the sample. The absorbance (OD value) is measured at 450 nm using a microplate reader, and sample concentration is calculated.
Species	Mouse
Description	This Monoamine Oxidase (MAO) ELISA Detection Kit provides a double-antibody sandwich ELISA method for the quantitative detection of MAO in mouse biological samples including serum, plasma, cell culture supernatant, and tissue homogenates. The kit uses a pre-coated 96-well microplate with anti-MAO capture antibody. The sandwich assay format utilizes HRP-labeled detection antibody, with TMB as the chromogenic substrate. The assay has a sensitivity of less than 1.0 U/L and a detection range. The kit is available in both 96-well and 48-well configurations and includes all necessary reagents for performing the assay.
Form	Liquid

Tel: 1-631-559-9269 1-516-512-3133
Email: info@creative-biomart.com Fax: 1-631-938-8127
45-1 Ramsey Road, Shirley, NY 11967, USA

Applications	Quantitative detection of monoamine oxidase (MAO) in mouse serum samples; Quantitative detection of MAO in mouse plasma samples (EDTA, citrate, or heparin anticoagulated); Quantitative detection of MAO in mouse cell culture supernatants; Quantitative detection of MAO in mouse tissue homogenates; Research on monoamine oxidase-related physiological and pathological processes in mouse models
Notes	This kit is for research use only and not for use in medical diagnosis or human experimentation. Store the kit at 2-8°C. Before use, equilibrate to room temperature for 20 minutes. Crystals may form in the concentrated wash buffer stored at 4°C; this is normal. Warm in a water bath to completely dissolve crystals before use. Unused plate strips should be immediately returned to the sealed bag and stored at low temperature in dry conditions. The S0 standard with concentration 0 can be regarded as negative control or blank. According to the instructions, samples have already been diluted 5-fold; the final result must be multiplied by 5 to obtain the actual sample concentration. Strictly follow the time, volume, and sequence of incubation operations as specified in the instructions. All liquid components should be thoroughly mixed before use. Read the OD values within 15 minutes after adding the stop solution. The standard curve can be plotted using linear regression in Excel. If samples are diluted, multiply by the corresponding dilution factor. Avoid repeated freeze-thaw cycles of samples. Ensure samples are uniformly and completely thawed at room temperature before testing.
Storage	Store at 2-8°C, protected from light and moisture. Valid for 6 months.
Shipping	Shipped with ice packs
Reaction conditions	Coating: Pre-coated with monoamine oxidase (MAO) antibody Incubation temperature: 37°C Incubation time: 60 min (after adding HRP-labeled detection antibody) Substrate incubation: 37°C, 15 min, protected from light Stop solution

Tel: 1-631-559-9269 1-516-512-3133
Email: info@creative-biomart.com Fax: 1-631-938-8127
45-1 Ramsey Road, Shirley, NY 11967, USA

	addition: 50 µL per well Measurement wavelength: 450 nm Reading time: Within 15 min after adding stop solution
Unit definition	Sensitivity: Less than 1.0 U/L Standard curve concentrations: 0, 25, 50, 100, 200, 400 U/L
Usage	Kit Components (96-well configuration): Pre-coated Microplate: 12 wells × 8 strips Standards (S0-S5): 0.3 mL × 6 tubes (Concentrations: 0, 25, 50, 100, 200, 400 U/L) Sample Diluent: 6 mL HRP-labeled Detection Antibody: 10 mL 20× Wash Buffer: 25 mL Substrate A: 6 mL Substrate B: 6 mL Stop Solution: 6 mL Plate Sealer: 2 sheets Instruction Manual: 1 copy Sealed Bag: 1 piece Kit Components (48-well configuration): Pre-coated Microplate: 12 wells × 4 strips Standards (S0-S5): 0.3 mL × 6 tubes (Concentrations: 0, 25, 50, 100, 200, 400 U/L) Sample Diluent: 3 mL HRP-labeled Detection Antibody: 5 mL 20× Wash Buffer: 15 mL Substrate A: 3 mL Substrate B: 3 mL Stop Solution: 3 mL Plate Sealer: 2 sheets Instruction Manual: 1 copy Sealed Bag: 1 piece Materials Required but Not Provided: Microplate reader (450 nm) High-precision pipettes and tips: 0.5-10 µL, 2-20 µL, 20-200 µL, 200-1000 µL 37°C incubator Distilled water Sample Collection, Processing, and Storage: 1. Serum: Use tubes without pyrogens and endotoxins. Avoid any cell stimulation during operation. After blood collection, centrifuge at 3000 rpm for 10 minutes to separate serum from red blood cells carefully and rapidly. 2. Plasma: Use EDTA, citrate, or heparin as anticoagulant. Centrifuge at 3000 rpm for 30 minutes and collect the supernatant. 3. Cell Culture Supernatant: Centrifuge at 3000 rpm for 10 minutes to remove particles and polymers. 4. Tissue Homogenate: Add appropriate amount of physiological saline to the tissue and homogenize. Centrifuge at 3000 rpm for 10 minutes and collect the supernatant. 5. Storage: If samples are not tested immediately after collection, aliquot according to single-use amount and store at -20°C. Avoid repeated freeze-thaw cycles. Thaw at room temperature and ensure samples are uniformly and completely thawed before testing. Reagent Preparation: Dilution of 20× Wash Buffer: Dilute with distilled water at 1:20 ratio, i.e., 1 part of 20×

Tel: 1-631-559-9269 1-516-512-3133
Email: info@creative-biomart.com Fax: 1-631-938-8127
45-1 Ramsey Road, Shirley, NY 11967, USA

Wash Buffer plus 19 parts of distilled water. Plate Washing Methods: 1. Manual Washing: Empty the wells thoroughly, fill each well with wash buffer, incubate for 1 min, then empty the wells again. Pat dry on absorbent paper. Repeat this washing procedure 5 times. 2. Automatic Plate Washer: Inject 350 µL of wash buffer per well, soak for 1 min, wash 5 times. Assay Procedure: 1. Remove the required number of strips from the aluminum foil bag after equilibrating at room temperature for 20 min. Return the remaining strips to the sealed bag and store at 4°C. 2. Set up standard wells and sample wells. Add 50 µL of different concentration standards to the standard wells. 3. For sample wells, first add 10 µL of sample, then add 40 µL of Sample Diluent. Do not add anything to blank wells. 4. Except for blank wells, add 100 µL of HRP-labeled detection antibody to each standard well and sample well. Seal the wells with plate sealers. Incubate in a 37°C water bath or incubator for 60 min. 5. Discard the liquid, pat dry on absorbent paper. Fill each well with wash buffer, incubate for 1 min, discard the wash buffer, and pat dry on absorbent paper. Repeat this washing procedure 5 times (or use an automatic plate washer). 6. Add 50 µL of Substrate A and 50 µL of Substrate B to each well. Incubate at 37°C protected from light for 15 min. 7. Add 50 µL of Stop Solution to each well. Read the OD values of each well at 450 nm wavelength within 15 min. Data Analysis: 1. Read the OD values of each well at 450 nm wavelength within 15 min using a microplate reader. 2. Plot the standard curve in Excel: Use standard concentration as the X-axis and corresponding OD value as the Y-axis to plot the standard linear regression curve. Calculate each sample concentration value according to the curve equation. 3. If samples are diluted, multiply by the corresponding dilution factor. Kit Performance: 1. Accuracy: The correlation coefficient R value of standard linear regression with expected concentration is greater than or equal to 0.9900. 2. Sensitivity: The minimum detection concentration is less than 1.0 U/L. 3. Specificity: No cross-reactivity with other soluble structural analogs. 4. Reproducibility: Both intra-assay and inter-assay coefficients of variation are less than 15%.

Tel: 1-631-559-9269 1-516-512-3133
Email: info@creative-biomart.com Fax: 1-631-938-8127
45-1 Ramsey Road, Shirley, NY 11967, USA