

Glutamic Oxaloacetic Transaminase Activity Assay Kit

Cat. No. Kit-0381 **Lot. No.** (See product label)

SPECIFICATION

Product Overview	Glutamic Oxaloacetic Transaminase Activity Assay Kit an amino group is transferred of from aspartate to alpha-ketoglutarate. The products of this reversible transamination reaction are oxaloacetate and glutamate. The glutamate is detected in a reaction that concomitantly converts a nearly colorless probe to color ($\lambda_{\text{max}} = 450$ nm). The kit provides a rapid, simple, sensitive and reliable test suitable as a high throughput activity assay of GOT with a detection limit of 10 mU per well.
Size	100 x 1 assay
Description	Glutamate-oxaloacetate transaminase (GOT), also known as Aspartate aminotransferase (AST) is a transaminase (EC 2.6.1.1) similar to the more liver specific alanine transaminase (ALT). Although commonly included clinically as part of a diagnostic liver function test, GOT has a broader clinical utility since it may also be elevated in diseases affecting other organs, such as the heart or muscles in myocardial infarction, also in acute pancreatitis, acute hemolytic anemia, severe burns, acute renal disease, musculoskeletal diseases and trauma. It catalyzes the reaction: Aspartate + alpha-Ketoglutarate = Oxaloacetate + Glutamate Diagnostically, it is almost always measured in units/liter (U/L).
Applications	Functional Studies
Storage	Store the kit at -20°C protected from light. Allow the Assay Buffer to warm to room temperature before use. Briefly centrifuge vials before opening. Read the entire protocol before performing the assay. GOT ENZYME MIX: Reconstitute with 220 μ L

 Tel: 1-631-559-9269 1-516-512-3133

 Email: info@creative-biomart.com  Fax: 1-631-938-8127

 45-1 Ramsey Road, Shirley, NY 11967, USA

dH₂O. Aliquot and store at -20°C. Use within two months. GOT DEVELOPER: Reconstitute with 820 µL dH₂O. Aliquot and store at -20°C. Use within two months. GOT SUBSTRATE: Reconstitute with 1.1 ml assay buffer. Store at -20°C. Use within two months. GOT POSITIVE CONTROL: Reconstitute with 100 µL dH₂O. Aliquot and store at -20°C. Use within two months. In the assay (optional), add 5 µL positive control and adjust the volume to 50 µL/well with Assay Buffer.

Kit Components	Components Identifier 100 tests Glutamate Standard (0.1M) Yellow 1 x 0.1ml Glutamic Oxaloacetic Transaminase Buffer WM 1 x 25ml Glutamic Oxaloacetic Transaminase Developer Red 1 vial Glutamic Oxaloacetic Transaminase Enzyme Mix Green 1 vial Glutamic Oxaloacetic Transaminase Positive Control Blue 1 vial Glutamic Oxaloacetic Transaminase Substrate Orange 1 vial
-----------------------	---

Detection method	Colorimetric
-------------------------	--------------

Compatible Sample Types	Cell culture supernatant, Urine, Serum, Plasma, Other biological fluids, Tissue Extracts
--------------------------------	--

Sensitivity	10 mU/well
--------------------	------------

 Tel: 1-631-559-9269 1-516-512-3133

 Email: info@creative-biomart.com  Fax: 1-631-938-8127

 45-1 Ramsey Road, Shirley, NY 11967, USA