

Urine Albumin-to-Creatinine Ratio Assay Kit

Cat. No. Kit-2161 **Lot. No.** (See product label)

SPECIFICATION

Product Overview	UACR Assay Kit provides a simple, sensitive, and high-throughput adaptable assay that detects albumin (detection range: 0.02- 2.5 mg/ml), creatinine (detection range: 0.002 -0.5 mg/ml) and urine albumin-to-creatinine ratio. The UACR ratio is determined in two steps: First, albumin is determined by using a probe (AB580) that specifically recognizes albumin (Ex/Em = 600/630 nm). Second, creatinine is converted to sarcosine via enzymatic reactions. Sarcosine is specifically oxidized generating a product that reacts with a probe producing a chromophore that can be detected at 570 nm.
Size	100 assays
Description	Urine Albumin-to-Creatinine Ratio (UACR) is one of the two markers used to determine chronic kidney disease (CKD). UACR is recommended to be measured on regular basis on people living with Type I and Type II diabetes. UACR is defined as the ratio between albumin (reported in mg/dl) and creatinine (reported in g/dl). This ratio estimates the amount of albumin excreted in urine during a 24 hr period. Albuminuria is diagnosed when UACR is greater than 30 mg albumin/g creatinine.
Applications	Estimation of albumin in biological samples Estimation of creatinine in biological samples Determination of UACR in mammalian urine samples
Storage	Store kit at -20°C, protected from light. Briefly spin small vials prior to opening. Read entire protocol before performing the assay. Creatinine Assay Buffer, Albumin Assay Buffer, and Albumin Diluent: Store at -20°C. Bring to room temperature (RT) before

 Tel: 1-631-559-9269 1-516-512-3133

 Email: info@creative-biomart.com  Fax: 1-631-938-8127

 45-1 Ramsey Road, Shirley, NY 11967, USA

use. Albumin Probe (AB580) and Creatinine Probe: Light sensitive. Store at -20°C. Bring to RT before use. Creatinase, Creatininase, and Creatinine Enzyme Mix: Reconstitute with 220 µl of Creatinine Assay Buffer. Aliquot and store at -20°C. Freeze/thaw should be limited to one time. Keep on ice during use. BSA Standard (2 mg/ml): Store at RT. Creatinine (10 µmol): Reconstitute with 115 µl of dH₂O to generate 10 g/µl Creatinine Standard. Dissolve completely. Store at -20°C. Use within 2 months.

Kit Components

Creatinine Assay Buffer: 25 ml
Albumin Assay Buffer: 7 ml
Albumin Diluent: 7 ml
Albumin Probe (AB580): 0.4 ml
Creatinine Probe: 0.2 ml
Creatinase: 1 vial
Creatininase: 1 vial
Creatinine Enzyme Mix: 1 vial
BSA Standard (2 mg/ml): 1 ml
Creatinine Standard (10 µmol): 1 vial

Detection method

Fluorescence (Ex/Em = 600/630 nm) & Absorbance (OD 570 nm)

Compatible Sample Types

• Albumin: urine, saliva, etc.
• Creatinine: urine, serum, etc.
• UACR: urine

Features & Benefits

• Simple & Sensitive
• High-Throughput adaptable
• Suitable for detecting albumin (detection range: 0.02- 2.5 mg/ml), creatinine (detection range: 0.002 -0.5 mg/ml) and urine albumin-to-creatinine ratio (UACR)

 Tel: 1-631-559-9269 1-516-512-3133

 Email: info@creative-biomart.com  Fax: 1-631-938-8127

 45-1 Ramsey Road, Shirley, NY 11967, USA